



## Fresenius Medical Care

**\*\*\* Important Prescribing Information \*\*\***

### **NaturaLyte Liquid and Granuflo Acid Concentrate**

### **Bicarbonate Alkalosis**

**DATE:** March 29, 2012

**SUBJECT:** Risk of Alkalosis with acetate containing dialysis acid concentrates

**PRODUCT CODES:** See Attached

Dear Unit Medical Director/Administrator/Director of Nursing/Home Therapies Manager/Customer,

Fresenius Medical Care North America (FMCNA) is issuing an urgent product notification involving the NaturaLyte Liquid and Granuflo powder product lines (Product Codes: See attached list). Both products contain acetate (NaturaLyte Liquid 4.0 mEq/L; Granuflo 8.0 mEq/L of acetate in the final dialysate); which in addition to bicarbonate, combine to yield the total prescribed buffer. Total buffer should be considered in addition to bicarbonate as part of writing the dialysis prescription.

Previous reports have identified an association between elevated pre-dialysis bicarbonate levels and an increased mortality risk.<sup>1,2,3,4</sup> Recent analyses performed using FMCNA hemodialysis (HD) patient safety data confirms that alkalosis is a significant risk factor associated with cardiopulmonary (CP) arrest in the dialysis unit, independent of and additive to the risk of CP arrest associated with pre-dialysis hypokalemia. A major cause of metabolic alkalosis in dialysis patients is inappropriately high dialysate total buffer concentration.

NaturaLyte Liquid contributes 4.0 mEq/L of acetate and Granuflo contributes 8.0 mEq/L of acetate to the final dialysate; which in addition to bicarbonate, combine to the total buffer that the patient receives from the dialysate. Acetate is also contained in the dialysis acid concentrates produced by other manufacturers. Since acetate is rapidly converted into bicarbonate by the liver, the bicarbonate prescription entered into the dialysis machine underestimates the total buffer that the patient receives from the dialysate by ~8 mEq/L with dialysate prepared from Granuflo (powder) or by ~4 mEq/L with dialysate prepared from NaturaLyte (liquid).



## Fresenius Medical Care

Fresenius recommends that clinicians exercise their best clinical judgment regarding the bicarbonate and total buffer base prescription for each patient. This includes individualizing dialysate prescriptions and reviewing them monthly with consideration of patient's pre-dialysis bicarbonate and dialysate total buffer.

Please complete and return the enclosed Reply Form, indicating receipt and understanding of this communication. If you have any additional questions, please contact Customer Service at 1-800-323-5188 or Medical Information at 1-855-616-2309.

Adverse reactions or quality problems experienced with the use of this product may be reported to the FDA's MedWatch Adverse Event Reporting by:

- Linking to the MedWatch website at [www.fda.gov/medwatch](http://www.fda.gov/medwatch)
- Calling 1-800-FDA-1088
- Faxing at 1-800-FDA-0178, or by
- Mailing to: MedWatch, HF-2, FDA, 5600 Fishers Lane, Rockville, MD 20852-9787

The FDA has been advised of this product notification

Sincerely,

Jose Diaz-Buxo, MD, FACP  
Senior Vice President  
Chief Medical and Regulatory Affairs Officer  
Fresenius Medical Care North America  
Renal Therapies Group

Enclosure:  
Reply Form

<sup>1</sup> Gennari FJ. Very low and high predialysis serum bicarbonate levels are risk factors for mortality: what are the Appropriate Interventions? *Semin Dial.* May-Jun;23(3):253-257 2010

<sup>2</sup> Wu DY, Shinaberger CS, Regidor DL, McAllister CJ, Kopple JD, Kalantar-Zadeh K: Association between serum bicarbonate and death in hemodialysis patients: is it better to be acidotic or alkalotic? *Clin J Am Soc Nephrol* 1:70-78, 2006

<sup>3</sup> Bommer J, Locatelli F, Satayathum S, Keen ML, Goodkin DA, Saito A, Akiba T, Port FK, Young EW: Association of predialysis serum bicarbonate levels with risk of mortality and hospitalization in the Dialysis Outcomes and Practice Patterns Study (DOPPS). *Am J Kidney Dis* 44:661-671 2004

<sup>4</sup> Lowrie EG, Lew NL: Death risk in hemodialysis patients: the predictive value of commonly measured variables and an evaluation of death rate differences between facilities. *Am J Kidney Dis* 15:458-482, 1990



# Fresenius Medical Care

## PRODUCT CODES

Product Name	Product Code
Manufacturer: Fresenius Medical Care North America	
NaturaLyte	08-0231-4
NaturaLyte	08-1001-0
NaturaLyte	08-1201-8
NaturaLyte	08-1231-3
NaturaLyte	08-1251-1
NaturaLyte	08-1301-4
NaturaLyte	08-2201-5
NaturaLyte	08-2231-2
NaturaLyte	08-2251-0
NaturaLyte	08-2301-3
NaturaLyte	08-2351-8
NaturaLyte	08-3201-4
NaturaLyte	08-3231-1
NaturaLyte	08-3251-9
NaturaLyte	08-3301-2
NaturaLyte	08-4123-1
NaturaLyte	08-4223-7
NaturaLyte	08-4225-1
NaturaLyte	08-4230-2
NaturaLyte	08-4231-0
NaturaLyte	08-4323-5
NaturaLyte	08-4325-1
GranuFlo	0FD1201-3B
GranuFlo	0FD1251-3B
GranuFlo	0FD2123-3B
GranuFlo	0FD2201-3B
GranuFlo	0FD2220-3B
GranuFlo	0FD2223-3B
GranuFlo	0FD2225-3B
GranuFlo	0FD2231-3B
GranuFlo	0FD2251-3B
GranuFlo	0FD2301-3B
GranuFlo	0FD2323-3B
GranuFlo	0FD2325-3B
GranuFlo	0FD3201-3B
GranuFlo	0FD3231-3B
GranuFlo	0FD3251-3B
GranuFlo	0FD3301-3B
Naturalyte	13-1251-1
Naturalyte	13-2201-5
Naturalyte	13-2231-2
Naturalyte	13-2251-0
Naturalyte	13-3231-1



## Fresenius Medical Care

<b>Product Name</b>	<b>Product Code</b>
Naturalyte	13-3251-9
Naturalyte	13-4123-1
Naturalyte	13-4220-1
Naturalyte	13-4225-1
Naturalyte	13-4325-1